

TEST REPORT

1. Applicant

Name : WISE CONTROL Inc.
Address : 199, Sanggal-dong, Giheung-gu, Yongin-si,
Gyeonggi-do, Korea

2. Products

Name : Pressure Gauge
Model/Type : P258
Manufacturer : WISE CONTROL Inc.
Remark : -

3. Test Standard/Method : IEC 60529:2001, KS C IEC 60529:2002

4. Test Results : IP67 (Refer to document)

5. Use of Report : For Quality Control.

6. Date of Application : 2010. 03. 15

7. Date of Issue : 2010. 03. 18

Tested by



Machinery Components Center
Chang-ho Kim

Approved by



Machinery Components Center
Leader Byung-woo Lee

The above test report is the accredited test results by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

Korea Testing Laboratory

222-13, Guro3-Dong Guro-Gu Seoul, 152-718, Korea.
<http://www.ktl.re.kr>

Tel. : +82-2-860-1531
Fax. : +82-2-860-1549