

# TEST REPORT



한국산업기술시험원  
Korea Testing Laboratory

Report No. : 16-062828-01-2

Page of Pages : ( 1 ) / ( 5 )



## 1. Client

Name : WISE CONTROL Inc.

Address : 2022, Deogyong-daero, Giheung-gu, Yongin-si, Gyeonggi-do, 17097, Korea

Date of Receipt : 2016. 10. 27

## 2. Use of Report : To verify IP grade to IEC 60529

## 3. Test Sample

Description : Low Differential Pressure Gauge

Manufacturer : WISE CONTROL Inc.

Model Name : P880

Serial Number : -

Remark : -

## 4. Date of Test : 2016. 11. 07. ~ 2016. 11. 08.

## 5. Test Standard/Method : IEC 60529: 2001

## 6. Testing Environment : Temperature : ( 24.3 ± 2.0 ) °C , Humidity : ( 49 ± 2 ) % R.H.

## 7. Test Results : Pass (IP67)

Note : 1. The test results contained apply only to the test sample(s) supplied by the client  
2. This test report shall not be reproduced in full or in part without approval of the KTL in advance.

|             |                                       |                                   |
|-------------|---------------------------------------|-----------------------------------|
| Affirmation | Tested by                             | Technical Manager                 |
|             | Name : Min Yeong-seung<br>(Signature) | Name : Kim Ji-hoon<br>(Signature) |

The above test report is the accredited test result by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA.

2016. 11. 09

**Korea Testing Laboratory**

Accredited by KOLAS, Republic of KOREA



87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel.+82-2-860-1540 Fax. +82-2-860-1549

FP202-03-03



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.