

TEST REPORT

Korea Safety Certification Co., Ltd.
1106, Cheoinseong-ro, Namsa-myeon, Cheoin-gu,
Yongin-si, Gyeonggi-do, Korea
Tel : 031-323-4905 Fax : 031-328-4905

Report No: 19-KR004
Page(1)/(7 Pages)



1. Client

- o Name: WISE CONTROL INC.
- o Address: 19F, 181, Gasan digital 1-ro, Geumcheon-gu, Seoul, Republic Korea

2. Sample Description (or Product/Model)

SMT2003 / Stainless steel 316L / Absolute pressure transmitter

3. Date of Test: 2019. 2. 18 ~ 2019. 2. 25

4. Test method used : KS C IEC 60529:2013, 13.1, 14.1

5. Test Results

Test Item	Test method used	Results
IP 6X	KS C IEC 60529:2013, 13.1	No dust ingress inside the enclosure.
IP X6	KS C IEC 60529:2013, 14.1	No water ingress inside the enclosure.
IP X7	KS C IEC 60529:2013, 14.1	No water ingress inside the enclosure.

The results shown in this test report only to the sample(s) tested unless otherwise stated.

Affirmation	Tested by Name : Kwangjin Ka (Signature)	Technical Manager Name : Hyonggil Kim (Signature)
-------------	---	--

The above test results are the result of the examination of the accredited field from KOLAS(Korea Laboratory Accreditation Scheme) who signed the ILAC(International Laboratory Accreditation Cooperation) MRA(Mutual Recognition Arrangement)

2019. 02. 26

Accredited from KOLAS Korea Safety Certification Co., Ltd.

